



DECLARATION OF PARTICIPATION IN INTERLABORATORY COMPARISON

GUM HARDNESS GAUGES SHORE A, SHORE D

ILC: K 07/17

**Organized in accordance with: requirements of international standard
ISO/IEC 17043: 2010 Conformity assessment. General requirements
for proficiency testing**

Participant: S.C.METROMAT S.R.L., Romania

Evaluation of participant:

Share of satisfactory results of participant in Interlaboratory comparison based on
criteria of normalized deviation ($E_n < 1$) is 100 %

In Banská Bystrica, on 13.11.2018


Tomáš Švantner
director of department for strategy and marketing

laboratoare participante	Valori E_n obtinute de Metromat
<p>4 laboratoare participante Romania (1) (Metromat)</p>	<p>Durimetru SHORE A $E_n = -0,66; -0,43; -0,57; -0,45; -0,40; -0,39; -0,56; -0,54; -0,5; -0,6$</p> <p style="text-align: right;">Media valorilor E_n: 0,51</p>
	<p>Durimetru SHORE D $E_n = -0,64; -0,59; -0,58; -0,24; -0,11; -0,15; -0,17; 0,03; -0,25; -0,45;$</p> <p style="text-align: right;">Media valorilor E_n: 0,321</p>