



C E R T I F I C A T E

ON PARTICIPATION IN THE INTERLABORATORY COMPARISON

No. 0318-OV-A90916-13

Subject of the ILC:

DIGITAL MULTIMETER DMM6

Carried out in the period of: May 2013 – January 2014

Identification of the ILC: **ILC 41-900-13**

Participant:

METROMAT SRL

Str. Ady Endre no. 44, 505600 Sacele - Brasov, Romania

Final statement

Following the Report No. 0318-ZV-A90900-13 on evaluation of the ILC measurement results the above mentioned participant

c o m p l i e s w i t h

the requirements of this interlaboratory comparison based on EN ISO/IEC 17043.

Date of issue: **10.02.2014**

RNDr. Simona Klenovská
Head of the Department for ILC



Ing. Jindřich Pošvář
Director for Legal Metrology

17 Laboratoare participante**Valori E_n obtinute de Metromat**

Cehia (10), Slovacia (4), Bulgaria,
Portugalia, **Romania (Metromat)**

DC-U: $E_n = -0,047; -0,070; 0,073; 0,089; 0,000; -0,094; -0,116;$
 $0,046; -0,239;$ **Media valorilor E_n : 0,086**

AC-U: $E_n = 0,513; 0,601; 0,155; 0,253; -0,036; -0,096; -0,126;$
 $0,062; 0,102; 0,064; 0,023;$ **Media valorilor E_n : 0,184**

DC-I: $E_n = 0,021; 0,062; -0,025; -0,023;$ **Media valorilor E_n : 0,032**

AC-I: $E_n = 0,162; 0,164; 0,010; -0,008;$ **Media valorilor E_n : 0,086**

R: $E_n = 0,091; 0,031; 0,068;$ **Media valorilor E_n : 0,063**