



CERTIFICATE

ON PARTICIPATION IN THE INTERLABORATORY COMPARISON

No. 0318-OV-A0420-14

Subject of the ILC:

TORQUE WRENCHES

Within the range: 12 Nm 200 Nm 400 Nm

Carried out in the period of: March 2014 – April 2015

ILC identification: **ILC 21-400-14**

Participant:

METROMAT SRL

Str. Ady Endre no. 44, 505600 Sacele - Brasov, ROMANIA


Final statement

Following the Report No. 0318-ZV-A0400-14 on evaluation of the ILC measurement results the above mentioned participant

c o m p l i e s w i t h

the requirements of this interlaboratory comparison based on EN ISO/IEC 17043.

Date of issue: **20.04.2015**


RNDr. Simona Klenovská
Head of the Department for ILC




Ing. František Staněk, PhD.
Director for Legal Metrology

21 Laboratoare participante

Valori E_n obtinute de Metromat

Cehia (7), Slovacia, Polonia, Spania,
Croatia (2), Estonia, Portugalia ,
Grecia, Cipru, Chile (3), Vietnam,
Romania (Metromat)

$E_n = 0,332; 0,449; 0,110; 0,271; -0,457; - 0,456; -0,204 ;$
Media valorilor $E_n : 0,325$

 Quality is a state of mind	ASOCIATIA LABORATOARELOR DIN ROMANIA - ROLAB Bucuresti, b-dul. Unirii, nr. 61, bl. F3, sc, TR2, et. 3, ap. 301, sector 3 office@rolab.ro, fax. 021.320.33.10	
	ROLAB-IC ACREDITAT RENAR ISO/CEI 17043-2010 Certificat de acreditare Nr. IC 008/13.10.2016	CERTIFICAT MRC ISO 9001 Certificat Nr. 279/2010

acreditat pentru
 FURNIZARE DE ÎNCERCĂRI
 DE COMPETENȚĂ



SR EN ISO/CEI 17043 2010
 CERTIFICAT DE ACREDITARE
 IC 008

CERTIFICAT Nr. 11.4/2017


Se certifica participarea **METROMAT SRL – Laborator de metrologie** la incercarile de competenta privind:

EVALUAREA PERFORMANTELOR LABORATOARELOR DE ETALONARE SISTEME DE INDICARE A FORȚEI ALE MAȘINILOR DE ÎNCERCAT MATERIALE

Cod SIC-ETA-FRT-01/2017

Rezultatele schemei de incercari de competenta sunt prezentate in **Raportul incercarii de competenta nr. 11/2017** emis la data de **31.07.2017**.

Codul laboratorului dvs. este:



Director ROLAB-IC
Steluta Duta
 Dr. ing. Steluta DUTA