



CERTIFICATE

ON PARTICIPATION IN THE INTERLABORATORY COMPARISON

No. 0318-OV-A0420-14

Subject of the ILC:

TORQUE WRENCHES

Within the range: 12 Nm 200 Nm 400 Nm

Carried out in the period of: March 2014 – April 2015

ILC identification: **ILC 21-400-14**

Participant:

METROMAT SRL

Str. Ady Endre no. 44, 505600 Sacele - Brasov, ROMANIA

Final statement

Following the Report No. 0318-ZV-A0400-14 on evaluation of the ILC measurement results the above mentioned participant

c o m p l i e s w i t h

the requirements of this interlaboratory comparison based on EN ISO/IEC 17043.

Date of issue: **20.04.2015**

RNDr. Simona Klenovská
Head of the Department for ILC



Ing. František Staněk, PhD.
Director for Legal Metrology

21 Laboratoare participante

Valori E_n obtinute de Metromat

Cehia (7), Slovacia, Polonia, Spania,
Croatia (2), Estonia, Portugalia ,
Grecia, Cipru, Chile (3), Vietnam,
Romania (Metromat)

$E_n = 0,332; 0,449; 0,110; 0,271; -0,457; - 0,456; -0,204 ;$
Media valorilor $E_n : 0,325$